HPLC Application

Increase Sample Through-Put (fig. 2)

	-	· • • • • • • • • • •	-	
Column:	Luna® 3 µm C18(2) 100 Å, LC Column 20 x 4 mm, Ea			
Dimensions:	20 x 4 mm ID			
Order No:	00M-4251-D0)		
Elution Type:	Isocratic			
Eluent A:	Water with 0.1% ammonium hydroxide, pH 10			
Eluent B:	Acetonitrile with 0.1% ammonium hydroxide, pH 10			
Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	80	20
Flow Rate:	2 mL/min			
Col. Temp.:	25 °C			
Detection:	UV-Vis AbsVariable Wave.(UV) @ 230 nm (ambient)			
Analyst Note:	SecurityGuard [™] Guard Cartridge System extends column lifetime.			
	- SecurityGuard Cartridges, C18 4 x 3.0mm, 10/Pk Part No.: AJ0-4287			
	- Holder Part No	.: KJ0-4282		
				2.0mm ID H = 74mAu 4.0mm ID H = 10mAu
\sim				



Products used in this application:



ANALYTES:

1 Pindolol

©2025 Phenomenex Inc. All rights reserved.

Phenomenex products are available worldwide.

www.phenomenex.com.cn

For more information contact your Phenomenex Representative at info@phenomenex.com

