HPLC Application

ID No.: **13625**

Diperodon

Column: N/A
Dimensions: N/A
Order No: N/A
Elution Type: N/A
Eluent A: Hexane/Dichloroethane/Ethanol-TFA 60:35:5 ethanol-TFA 20:1

Gradient Profile:	**Step No.**	**Time (min)**	**Pct A**
	1	0	100

Flow Rate: 1 mL/min
Col. Temp.: ambient
Detection: UV-Vis Abs.-Variable Wave.(UV) @ 242 nm (ambient)



ANALYTES:

1 Diperodon
Retention Time: 18.3 min

2 Diperodon
Retention Time: 18.85 min

©2025 Phenomenex Inc. All rights reserved.