## **HPLC Application**

ID No.: 16194



## Jupiter® 300 C4 - Comparison of Mobile Phase Conditions

Jupiter® 5 µm C4 300 Å, LC Column 150 x 4.6 mm, Ea

150 x 4.6 mm ID **Dimensions:** Order No: 00F-4167-E0 Elution Type: Gradient

Eluent A: 0.1% TFA in Water

90% ACN 0.09% TFA in Water Eluent B:

Step No. Gradient Time (min) Pct B Pct A **Profile:** 0 80 20 2 25 45 55

Flow Rate: 1 mL/min 45 °C Col. Temp.:

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 214 nm (ambient)


Reversed phase runs of PEGylated carbonic anhydrase and PEGylated ribonuclease A using different organic mobile pha 3 con Analyst Note:

colution of different PEGylated forms.



Products used in this application:





## **ANALYTES:**

1 PEGylated Carbonic anhydrase or PEGylated Ribonucluease A

@2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at info@phenomenex.com

