

Poly dT 12-18 standard on Oligo-RP (modifying TEA/HFIP Ratio)

Column: Clarity® 3 µm Oligo-RP, LC Column 150 x 2 mm, Ea

Dimensions: 150 x 2 mm ID

Order No: 00F-4441-B0

Elution Type: Gradient

Eluent A: TEAA/HFIP/2%MeOH in Water

Eluent B: Acetonitrile


Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	100	0
	2	10	90	10
	3	30	70	30

Flow Rate: 0.2 mL/min

Col. Temp.: ambient

Detection: Mass Spectrometer (MS) @ 400 amu (ambient)

Analyst Note: Different ratios of HFIP to TEA were investigated using the poly dT 12-18 standard mixture. HFIP was fixed at 200 mM and different levels of TEA were used between 1 and 20 mM (4mm-red trace {pH 7.6}, 8mm-blue trace{pH 8.0}, and 14mm -black trace{pH8.3}). Optimal retention, MS



ANALYTES:

- 1 Poly dT 12-18 Oligonucleotide

Retention Time: 1 min



Products used in this application:

