HPLC Application

ID No.: 19760



Testosterone from Male Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 μm C18

Kinetex® 1.7 μ m C18 100 Å, LC Column 30 x 2.1 mm, Ea

30 x 2.1 mm ID **Dimensions:** Order No: 00A-4475-AN Elution Type: Gradient

Eluent A: 0.1% Formic Acid +1 mM Amm Formate in Water 0.1% Formic Acid +1 mM Amm Formate in ACN Fluent B:

Liuent D.	0.1 /0 Formic Acid 11 min Amin Formace in Acid					
Gradient	Step No.	Time (min)	Pct A	Pct B		
Profile:	1	0	90	10		
	2	2.5	10	90		
	3	3.5	10	90		
	4	3.6	90	10		

Kinetex^e

Products used in this application:




Flow Rate: 0.4 mL/min Col. Temp.: ambient

Detection: Mass Spectrometer (MS) @ amu (ambient)

Detector Info: <a target="_blank"

rts/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIEX< 19760



ANALYTES:

Testosterone

Retention Time: 2.62 min 2 Testosterone-d3 Retention Time: 2.61 min

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at info@phenomenex.com



Phenomenex products are available worldwide. www.phenomenex.com.cn

info@phenomenex.com

Sample Preparation Details

for HPLC Application ID No.: 19760



Testosterone from Male Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 μm C18

PRODUCT DESCRIPTION:

Strata[™]-X-A 33 µm Polymeric Strong Anion, 30 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S123-TBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE:

Note: The solvent volumes shown below are for a 30 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:		-
Load:		-
Wash:		_
Dry:		
Dry for	5 min under high vacuum	-
Elute:		-
Final Prep	and Analysis:	_
	ng evaporation of elution solvent $@$ 50-55 C under gentle nitrogen stream; uL 25% hydroxylamine solution and heat at 60-65 C for 5-10 min, then add	200 uL 5%
Inject:	0 μL on HPLC Mass Spectrometer (MS) @ amu (ambient)	

ANALYTES:	Spiked Conc.	Log P	рКа	% Rec	%RSC
	(ng/mL)				(n=0)
1 Testosterone	0				
2 Testosterone-d3	0.5				

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at info@phenomenex.com



Phenomenex products are available worldwide.