



Aeris Batch Reproducibility #4: ABN7 on Aeris PEPTIDE 1.7u XB-C18 50x4.6mm**Column:** Aeris[™] 1.7 μ m PEPTIDE XB-C18 100 Å, LC Column 50 x 2.1 mm, Ea**Dimensions:** 50 x 2.1 mm ID**Order No:** 00B-4506-AN**Elution Type:** Gradient**Eluent A:** 0.1% Formic Acid in Water**Eluent B:** 0.1% Formic Acid in AcCN

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	95	5
	2	4	5	95
	3	6	95	5

Flow Rate: 0.384 mL/min**Col. Temp.:** 30 °C**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)**Analyst Note:** Actual QC testing is done on a 50 x 4.6 mm running at 1.85 mL/min. Column is not an active PN so column was entered as 100 x 2.1 mm running at 0.384 mL/min for those that would want to replicate the experiment on a commercially available column**Products used in this application:****ANALYTES:****1** ABN 7 Mixture