HPLC Application

ID No.: 20771



Kinetex[®]

Heterocyclic Amines by LCMS using Kinetex 5u C18 50x2.1mm

Kinetex® 5µm C18 100 Å, LC Column 50 x 2.1 mm, Ea

50 x 2.1 mm ID **Dimensions:** Order No: 00B-4601-AN **Elution Type:** Gradient

Eluent A: 30 mM Ammonium formate

Eluent B: Acetonitrile

| Gradient | Step No. | Time (min) | Pct A | Pct B |
|----------|----------|------------|-------|-------|
| Profile: | 1 | 0 | 95 | 5 |
| | 2 | 5 | 40 | 60 |
| | 3 | 5.1 | 0 | 100 |
| | 4 | 6 | 0 | 100 |
| | 5 | 7 | 95 | 5 |


Products used in this application:


500 μL/min Flow Rate: Col. Temp.: ambient

Detection: Tandem Mass Spec (MS-MS) @ (ambient)

Detector Info: <a target="_blank" hrsf="https://sciav

/products/mass-spactromaters?utm_campaign=2010%20application%20aparch&utm_source=phenomenex&utm_medium=referral">SCIEX< 20771







HPLC Application ID No.: **20771**



Heterocyclic Amines by LCMS using Kinetex 5u C18 50x2.1mm **ANALYTES:**

- DMIP 1
- 2 IQx
- 3 ΙQ
- 4 8-MeIQ
- 4,8-DiMe-IQx 5
- 7,8-DiMe-IQx 6
- 7 Trp-P-2
- Harman
- 9 Norharman
- **10** PhIP
- **11** AaC
- 12 MeAaC
- **13** MeIQ

Sample Preparation Details

for **HPLC** Application ID No.: **20771**



Heterocyclic Amines by LCMS using Kinetex 5u C18 50x2.1mm

PRODUCT DESCRIPTION:

Strata[™]-X-C 33 µm Polymeric Strong Cation, 60 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S029-UBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE:

Note: The solvent volumes shown below are for a 60 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

| Condition: |
|--|
| Load: |
| Wash: |
| Dry: |
| Elute: |
| Final Prep and Analysis: |
| Inject: 2 ul. on HPLC Tandem Mass Spec (MS-MS) @ (ambient) |

| ANAL | YTES: | Spiked Conc. (ng/mL) | Log P | pKa | % Rec | %RSC (n=0) |
|------|--------------|-------------------------|-------|-----|-------|---------------|
| 1 | DMIP | 0 | | | | |
| 2 | IQx | 0 | | | | |
| 3 | IQ | 0 | | | | |
| 4 | 8-MeIQ | 0 | | | | |
| 5 | 4,8-DiMe-IQx | 0 | | | | |
| 6 | 7,8-DiMe-IQx | 0 | | | | |
| 7 | Trp-P-2 | 0 | | | | |
| 8 | Harman | 0 | | | | |
| 9 | Norharman | 0 | | | | |
| 10 | PhIP | 0 | | | | |
| 11 | AaC | 0 | | | | |
| 12 | MeAaC | 0 | | | | |
| 13 | MeIQ | 0 | | | | |

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at info@phenomenex.com



Phenomenex products are available worldwide.

www.phenomenex.com.cn info@phenomenex.com