

Calibration Curves for (R)-Amlodipine and (S)-Amlodipine

Column: Lux® 3 µm Cellulose-4, LC Column 150 x 2 mm, Ea
Dimensions: 150 x 2 mm ID
Order No: 00F-4490-B0
Elution Type: Gradient
Eluent A: 0.05% Ethanolamine in Acetonitrile and Isopropyl Alcohol (96:4 v/v)
Gradient


Step No.	Time (min)	Pct A
1	0	100
2	4	100

Profile:

Flow Rate: 0.3 mL/min
Col. Temp.: 25 °C
Detection: LC/MS/MS @ 0.0000000000 (ambient)



Products used in this application:



ANALYTES:

- 1 (R)-Amlodipine
Retention Time: 2.294 min
- 2 (S)-Amlodipine
Retention Time: 2.295 min

Sample Preparation Details

for HPLC Application ID No.: 25532

Calibration Curves for (R)-Amlodipine and (S)-Amlodipine

PRODUCT DESCRIPTION:

StrataTM-X 33 μ m Polymeric Reversed Phase, 30 mg / 1 mL, Tubes , 100/Pk

Order No.: 8B-S100-TAK

SOLID PHASE EXTRACTION (SPE) PROCEDURE:

Note: The solvent volumes shown below are for a bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:

Load:

5 μ L of internal standard working solution was added to 100 μ L spiked human plasma (spiked at 0.050, 0.150, 20.0, 40.0, and 50.0 ng/mL) and mixed, 500 μ L of 0.2% (v/v) Ethanolamine in water was added to sample and vortexed.

Wash:

Dry:

Elute:

Final Prep and Analysis:

Inject: 10 μ L on HPLC LC/MS/MS @ 0.0000000000 (ambient)

ANALYTES:	Spiked Conc. (ng/mL)	Log P	pKa	% Rec	%RSC (n=0)
1 (R)-Amlodipine	0				
2 (S)-Amlodipine	0				

Note: This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals.
Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at info@phenomenex.com



Phenomenex products are available worldwide.

www.phenomenex.com.cn

info@phenomenex.com